
Midmark M11 - v1.0.5
Total Cycles: 531

Name: RITTER2
Sterilizer ID: V2313228
11/09/2021 03:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.7 F 0.0
2:00 186.9 F 4.3
4:00 202.2 F 9.5
6:00 217.1 F 7.2
8:00 231.2 F 10.9
10:00 233.5 F 8.7
12:00 238.8 F 10.2
14:00 248.3 F 14.9
16:00 261.0 F 22.1
18:00 272.6 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.0 F 29.3
2:00 272.3 F 29.7
3:00 271.5 F 29.2
4:00 272.2 F 29.5

Min 271.4 F 29.0
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.6 F 0.0
5:00 233.0 F 0.0
10:00 235.8 F 0.0
15:00 226.8 F 0.0
20:00 225.5 F 0.0
25:00 230.1 F 0.0
30:00 227.4 F 0.0
Drying Complete

FILLING: 2:37
HEATING: 18:22
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:08
CYCLE COMPLETE

11/09/2021 04:35 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____