

-----  
Midmark M11 - v1.0.0  
Total Cycles: 804

Name: RITTER 1  
Sterilizer ID: V2178198  
04/26/2021 06:34 AM

BEGIN UNWRAPPED CYCLE  
Temp: 270 Degrees F  
Time: 3 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.7 F 0.0  
2:00 72.1 F 0.0  
4:00 78.6 F 0.0  
6:00 111.1 F 0.1  
8:00 145.8 F 0.2  
10:00 184.9 F 0.3  
12:00 201.3 F 3.5  
14:00 224.7 F 6.2  
16:00 229.5 F 9.4  
18:00 236.9 F 13.2  
20:00 235.0 F 8.7  
22:00 243.8 F 11.9  
24:00 243.4 F 11.9  
26:00 240.9 F 10.5  
28:00 254.6 F 17.4  
30:00 266.1 F 24.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 27.9  
1:00 272.4 F 28.3  
2:00 272.3 F 28.5  
3:00 271.9 F 28.0

Min 271.5 F 27.7  
Max 272.8 F 28.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 214.0 F 0.0  
5:00 226.7 F 0.0  
10:00 219.9 F 0.0  
15:00 223.3 F 0.0  
20:00 185.1 F 0.0  
25:00 196.1 F 0.0  
30:00 186.3 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 31:36  
STERILIZING: 3:00  
VENTING: 2:60  
DRYING: 30:00  
TOTAL CYCLE: 01:08:24  
CYCLE COMPLETE

04/26/2021 07:42 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01