

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3335

Name: RITTER 1  
Sterilizer ID: V2178198  
06/05/2023 12:19 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 156.6 F 0.0  
2:00 157.7 F 0.0  
4:00 182.1 F 1.8  
6:00 203.2 F 6.2  
8:00 218.1 F 11.1  
10:00 224.3 F 6.3  
12:00 237.2 F 11.6  
14:00 235.5 F 8.9  
16:00 240.0 F 10.0  
18:00 247.5 F 13.9  
20:00 259.6 F 20.6  
22:00 270.6 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 272.5 F 28.9  
2:00 271.5 F 28.2  
3:00 272.6 F 29.0  
4:00 271.9 F 28.3

Min 271.5 F 28.1  
Max 272.6 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.8 F 0.0  
5:00 221.2 F 0.0  
10:00 222.1 F 0.0  
15:00 222.5 F 0.0  
20:00 218.9 F 0.0  
25:00 214.7 F 0.0  
30:00 210.5 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 22:48  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 01:00:59  
CYCLE COMPLETE

06/05/2023 01:19 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_