

-----  
Midmark M11 - v1.0.0  
Total Cycles: 833

Name: RITTER 1  
Sterilizer ID: V2178198  
05/20/2021 02:50 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 160.2 F 0.0  
2:00 157.8 F 1.8  
4:00 168.8 F 5.1  
6:00 192.1 F 9.8  
8:00 218.9 F 6.3  
10:00 226.8 F 10.4  
12:00 229.0 F 7.0  
14:00 243.7 F 13.5  
16:00 243.4 F 12.1  
18:00 256.4 F 19.0  
20:00 268.1 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.9  
1:00 272.1 F 28.8  
2:00 272.5 F 29.0  
3:00 272.6 F 29.3  
4:00 272.4 F 29.5

Min 271.7 F 28.8  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.7 F 0.0  
5:00 230.0 F 0.0  
10:00 225.7 F 0.0  
15:00 237.7 F 0.0  
20:00 225.4 F 0.0  
25:00 226.6 F 0.0  
Dry Cycle Aborted

FILLING: 1:38  
HEATING: 21:43  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 28:23  
TOTAL CYCLE: 00:58:51

05/20/2021 03:48 PM  
ACCEPTED:  
REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_