

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4019

Name: RITTER2  
Sterilizer ID: V2313228  
10/10/2024 12:46 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 168.0 F 0.0  
2:00 167.6 F 2.6  
4:00 182.8 F 6.9  
6:00 208.5 F 11.9  
8:00 219.2 F 7.0  
10:00 234.3 F 12.6  
12:00 236.1 F 10.4  
14:00 235.7 F 9.1  
16:00 250.7 F 16.2  
18:00 263.0 F 23.4  
20:00 271.7 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.3  
1:00 272.2 F 29.3  
2:00 272.4 F 29.8  
3:00 271.5 F 29.6  
4:00 272.0 F 29.4

Min 271.3 F 29.1  
Max 272.7 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.1 F 0.0  
5:00 212.7 F 0.0  
10:00 214.5 F 0.0  
15:00 210.3 F 0.0  
20:00 194.3 F 0.0  
25:00 187.1 F 0.0  
30:00 182.9 F 0.0  
Drying Complete

FILLING: 1:38  
HEATING: 20:05  
STERILIZING: 4:00  
VENTING: 3:30  
DRYING: 30:00  
TOTAL CYCLE: 00:59:14  
CYCLE COMPLETE

10/10/2024 01:45 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_