
Midmark M11 - v1.0.5
Total Cycles: 2470

Name: RITTER2
Sterilizer ID: V2313228
06/06/2023 10:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.7 F 0.0
2:00 184.6 F 3.3
4:00 195.4 F 8.4
6:00 216.5 F 10.3
8:00 228.1 F 9.7
10:00 233.7 F 8.8
12:00 243.7 F 13.6
14:00 243.8 F 12.7
16:00 256.5 F 19.5
18:00 267.7 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.4 F 29.4
2:00 272.1 F 29.9
3:00 272.0 F 29.1
4:00 272.4 F 29.8

Min 271.4 F 29.0
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.1 F 0.0
5:00 231.0 F 0.0
10:00 240.2 F 0.0
15:00 229.9 F 0.0
20:00 226.9 F 0.0
25:00 228.0 F 0.0
30:00 226.0 F 0.0
Drying Complete

FILLING: 1:58
HEATING: 19:19
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:31
CYCLE COMPLETE

06/06/2023 11:07 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____