

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 000503  
12:35 07/05/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:42  
PRESSURIZING 3:43  
STERILIZING 4:56  
136.0 C 230kPa 4:56  
136.3 C 232kPa 5:26  
136.2 C 231kPa 5:56  
136.4 C 233kPa 6:26  
136.4 C 233kPa 6:56  
136.2 C 232kPa 7:26  
136.4 C 233kPa 7:56  
136.4 C 233kPa 8:26  
136.4 C 233kPa 8:26  
VENTING 8:26  
AIR DRYING 9:23  
CYCLE COMPLETE 24:23  
-----

Digital Signature #  
663B0EBEB4799C75  
-----