
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001962
9:39 16/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:26
PRESSURIZING 2:32
STERILIZING 3:29
136.0 C 221kPa 3:29
136.4 C 225kPa 3:59
136.2 C 225kPa 4:29
136.2 C 224kPa 4:59
136.4 C 226kPa 5:29
136.3 C 223kPa 5:59
136.4 C 226kPa 6:29
136.2 C 224kPa 6:59
136.2 C 224kPa 6:59
VENTING 6:59
AIR DRYING 7:47
CYCLE COMPLETE 11:20

Digital Signature #
8C2194EEE69774C5
