
Midmark M11 - v1.0.5
Total Cycles: 2176

Name: RITTER2
Sterilizer ID: V2313228
03/27/2023 09:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 139.2 F 0.0
2:00 139.3 F 1.8
4:00 149.0 F 4.4
6:00 175.3 F 8.6
8:00 202.0 F 12.2
10:00 213.1 F 7.7
12:00 227.9 F 13.1
14:00 234.2 F 10.1
16:00 235.3 F 8.4
18:00 247.8 F 15.1
20:00 260.0 F 21.9
22:00 271.3 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.5
1:00 272.4 F 29.8
2:00 271.9 F 30.1
3:00 271.9 F 29.6
4:00 272.1 F 30.2

Min 271.4 F 29.4
Max 272.6 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.9 F 0.0
5:00 224.1 F 0.0
10:00 221.7 F 0.0
15:00 220.9 F 0.0
20:00 222.1 F 0.0
25:00 223.9 F 0.0
30:00 223.1 F 0.0
Drying Complete

FILLING: 1:00
HEATING: 22:38
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:00:52
CYCLE COMPLETE

03/27/2023 10:14 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____