

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2439

Name: RITTER 1  
Sterilizer ID: V2178198  
10/17/2022 10:00 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 197.0 F 0.0  
2:00 201.0 F 4.0  
4:00 206.3 F 8.6  
6:00 215.0 F 11.5  
8:00 227.6 F 8.2  
10:00 237.9 F 13.5  
12:00 238.2 F 10.7  
14:00 237.3 F 8.7  
16:00 250.0 F 15.5  
18:00 261.8 F 22.2  
20:00 272.6 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.6  
1:00 272.0 F 29.1  
2:00 272.3 F 29.0  
3:00 271.4 F 28.5  
4:00 272.6 F 29.4  
  
Min 271.4 F 28.4  
Max 272.6 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.0 F 0.0  
5:00 224.7 F 0.0  
10:00 231.1 F 0.0  
15:00 227.8 F 0.0  
20:00 225.0 F 0.0  
25:00 224.1 F 0.0  
30:00 222.7 F 0.0  
Drying Complete

FILLING: 1:37  
HEATING: 20:18  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:07  
CYCLE COMPLETE

10/17/2022 10:59 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_