
Midmark M11 - v1.0.5
Total Cycles: 1174

Name: RITTER2
Sterilizer ID: V2313228
06/08/2022 11:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 135.5 F 0.0
2:00 137.1 F 0.9
4:00 152.3 F 3.9
6:00 179.4 F 8.1
8:00 200.4 F 13.0
10:00 218.3 F 6.7
12:00 232.0 F 11.5
14:00 232.5 F 8.5
16:00 240.9 F 11.3
18:00 245.0 F 13.3
20:00 257.2 F 19.9
22:00 268.2 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.2
1:00 272.6 F 29.9
2:00 272.0 F 29.3
3:00 271.6 F 29.5
4:00 272.9 F 29.9

Min 271.3 F 29.0
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.0 F 0.0
5:00 212.7 F 0.0
10:00 228.6 F 0.0
15:00 226.7 F 0.0
20:00 225.5 F 0.0
25:00 223.6 F 0.0
30:00 222.9 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 23:19
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:01:24
CYCLE COMPLETE

06/08/2022 12:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____