
Midmark M11 - v1.0.5
Total Cycles: 2696

Name: RITTER2
Sterilizer ID: V2313228
07/25/2023 10:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.9 F 0.0
2:00 171.1 F 0.7
4:00 180.5 F 4.2
6:00 197.5 F 8.1
8:00 223.5 F 12.3
10:00 225.9 F 7.0
12:00 236.1 F 11.6
14:00 234.3 F 8.7
16:00 241.3 F 11.2
18:00 246.0 F 13.6
20:00 257.4 F 19.8
22:00 267.8 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.9
1:00 272.5 F 29.5
2:00 271.5 F 28.8
3:00 272.2 F 29.7
4:00 272.1 F 29.2

Min 271.2 F 28.7
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.3 F 0.0
5:00 211.8 F 0.0
10:00 216.1 F 0.0
15:00 214.6 F 0.0
20:00 209.8 F 0.0
25:00 207.2 F 0.0
30:00 204.2 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 23:24
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:02:13
CYCLE COMPLETE

07/25/2023 11:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____