
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
1.0uS / 0.6ppm
CYCLE NUMBER 001547
10:57 11/01/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 5:02
PRESSURIZING 7:00
STERILIZING 8:45
136.0 C 220kPa 8:45
136.3 C 217kPa 9:15
136.4 C 222kPa 9:45
136.3 C 220kPa 10:15
135.6 C 217kPa 10:45
136.3 C 221kPa 11:15
136.1 C 221kPa 11:45
136.2 C 221kPa 12:15
VENTING 12:15
AIR DRYING 13:20
CYCLE COMPLETE 35:55

Digital Signature #
3AFB313D4B0DBD38
