
Midmark M11 - v1.0.0
Total Cycles: 2603

Name: RITTER 1
Sterilizer ID: V2178198
11/23/2022 11:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.7 F 0.0
2:00 186.3 F 4.2
4:00 195.8 F 9.3
6:00 220.4 F 7.1
8:00 228.6 F 9.8
10:00 233.7 F 8.1
12:00 243.0 F 13.3
14:00 242.6 F 12.1
16:00 255.9 F 18.9
18:00 267.6 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 272.6 F 29.4
2:00 272.0 F 28.8
3:00 271.8 F 29.0
4:00 272.4 F 29.5

Min 271.5 F 28.6
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.7 F 0.0
5:00 206.2 F 0.0
10:00 220.9 F 0.0
15:00 223.0 F 0.0
20:00 223.3 F 0.0
25:00 222.9 F 0.0
30:00 222.7 F 0.0
Drying Complete

FILLING: 1:47
HEATING: 19:20
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:21
CYCLE COMPLETE

11/23/2022 12:55 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____