
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003056
11:36 01/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:38
PRESSURIZING 2:43
STERILIZING 3:42
136.4 C 220kPa 3:42
136.1 C 222kPa 4:12
136.2 C 221kPa 4:42
136.4 C 221kPa 5:12
136.4 C 222kPa 5:42
136.4 C 219kPa 6:12
136.3 C 222kPa 6:42
136.4 C 225kPa 7:12
136.4 C 225kPa 7:12
VENTING 7:12
AIR DRYING 8:04
CYCLE COMPLETE 8:15

Digital Signature #
06C4F9540DAB44B9
