
Midmark M11 - v1.0.0
Total Cycles: 1316

Name: RITTER 1
Sterilizer ID: V2178198
10/28/2021 01:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.6 F 0.0
2:00 152.3 F 2.0
4:00 172.6 F 5.3
6:00 195.8 F 9.8
8:00 219.9 F 6.0
10:00 230.8 F 8.6
12:00 240.4 F 12.2
14:00 240.3 F 10.7
16:00 235.7 F 8.6
18:00 250.8 F 15.6
20:00 262.8 F 22.4
22:00 272.3 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.1 F 28.4
2:00 272.4 F 28.9
3:00 271.8 F 28.6
4:00 271.8 F 28.4

Min 271.4 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.1 F 0.0
5:00 225.5 F 0.0
10:00 232.7 F 0.0
15:00 219.2 F 0.0
20:00 219.7 F 0.0
25:00 219.4 F 0.0
30:00 211.9 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 22:10
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:00:44
CYCLE COMPLETE

10/28/2021 02:06 PM
ACCEPTED: X
REJECTED:
Operator: U02