
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000683
8:01 15/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:19
PRESSURIZING 2:26
STERILIZING 3:28
136.0 C 227kPa 3:28
136.4 C 233kPa 3:58
136.3 C 232kPa 4:28
136.1 C 230kPa 4:58
136.4 C 232kPa 5:28
136.1 C 230kPa 5:58
136.3 C 231kPa 6:28
136.4 C 232kPa 6:58
136.4 C 232kPa 6:58
VENTING 6:58
AIR DRYING 7:51
CYCLE COMPLETE 7:56

Digital Signature #
72D388A91CE735D4
