
Midmark M11 - v1.0.0
Total Cycles: 5268

Name: RITTER 1
Sterilizer ID: V2178198
01/23/2025 01:57 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.2 F 0.0
2:00 166.4 F 3.2
4:00 181.3 F 6.7
6:00 210.3 F 10.6
8:00 218.2 F 4.0
10:00 223.2 F 7.7
12:00 233.8 F 12.2
14:00 234.8 F 8.5
16:00 244.4 F 12.2
18:00 244.4 F 12.4
20:00 256.2 F 18.5
22:00 266.6 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.3 F 29.1
2:00 272.6 F 28.8
3:00 271.5 F 28.3
4:00 271.3 F 28.5

Min 271.1 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.0 F 0.0
5:00 215.8 F 0.0
10:00 197.4 F 0.0
15:00 197.3 F 0.0
20:00 205.2 F 0.0
25:00 237.1 F 0.0
30:00 225.4 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 23:39
STERILIZING: 4:00
VENTING: 3:56
DRYING: 30:00
TOTAL CYCLE: 01:02:37
CYCLE COMPLETE

01/23/2025 02:59 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____