
Midmark M11 - v1.0.5
Total Cycles: 1170

Name: RITTER2
Sterilizer ID: V2313228
06/07/2022 03:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.6 F 0.0
2:00 170.0 F 3.0
4:00 174.5 F 7.2
6:00 197.0 F 12.3
8:00 218.1 F 6.9
10:00 231.3 F 12.0
12:00 232.6 F 9.1
14:00 238.2 F 10.1
16:00 245.6 F 13.9
18:00 257.4 F 20.3
20:00 268.5 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.4
1:00 272.7 F 30.2
2:00 272.4 F 29.8
3:00 271.5 F 29.3
4:00 271.9 F 30.0

Min 271.4 F 29.2
Max 272.9 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.2 F 0.0
5:00 215.4 F 0.0
10:00 227.0 F 0.0
15:00 224.8 F 0.0
20:00 221.9 F 0.0
25:00 226.6 F 0.0
30:00 227.3 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 21:19
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:57
CYCLE COMPLETE

06/07/2022 04:18 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____