

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2385

Name: RITTER 1  
Sterilizer ID: V2178198  
10/03/2022 09:33 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 218.0 F 0.0  
2:00 211.5 F 3.5  
4:00 218.0 F 8.6  
6:00 226.8 F 8.0  
8:00 235.1 F 10.2  
10:00 237.2 F 7.8  
12:00 243.1 F 11.6  
14:00 246.1 F 13.1  
16:00 258.9 F 20.1  
18:00 270.3 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.3  
1:00 272.4 F 28.7  
2:00 272.0 F 28.8  
3:00 272.1 F 28.6  
4:00 272.0 F 28.9  
  
Min 271.5 F 28.2  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 220.1 F 0.0  
5:00 230.3 F 0.0  
10:00 240.4 F 0.0  
15:00 231.0 F 0.0  
20:00 228.6 F 0.0  
25:00 229.5 F 0.0  
30:00 229.6 F 0.0  
Drying Complete

FILLING: 2:05  
HEATING: 18:50  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 00:58:02  
CYCLE COMPLETE

10/03/2022 10:31 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_