

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001785  
11:37 16/06/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:31  
PRESSURIZING 2:32  
STERILIZING 3:31  
136.4 C 224kPa 3:31  
136.2 C 222kPa 4:01  
136.3 C 223kPa 4:31  
136.3 C 222kPa 5:01  
136.3 C 223kPa 5:31  
136.4 C 224kPa 6:01  
136.1 C 223kPa 6:31  
136.4 C 222kPa 7:01  
136.4 C 222kPa 7:01  
VENTING 7:02  
AIR DRYING 7:54  
CYCLE COMPLETE 8:01

-----  
Digital Signature #  
65A46313F050F8F3  
-----