
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.9uS / 0.5ppm
CYCLE NUMBER 001476
13:24 10/06/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:36
PRESSURIZING 6:17
STERILIZING 7:48
136.0 C 220kPa 7:48
136.2 C 221kPa 8:18
136.4 C 222kPa 8:48
136.1 C 219kPa 9:18
136.5 C 222kPa 9:48
136.4 C 220kPa 10:18
136.5 C 222kPa 10:48
136.4 C 222kPa 11:18
VENTING 11:18
AIR DRYING 12:18
CYCLE COMPLETE 16:49

Digital Signature #
D470A4242F9FCA15
