
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001944
8:51 12/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:21
PRESSURIZING 2:26
STERILIZING 3:34
136.0 C 220kPa 3:34
136.4 C 225kPa 4:04
136.4 C 226kPa 4:34
136.4 C 222kPa 5:04
136.3 C 224kPa 5:34
136.4 C 226kPa 6:04
136.3 C 225kPa 6:34
136.2 C 224kPa 7:04
136.2 C 224kPa 7:04
VENTING 7:05
AIR DRYING 7:57
CYCLE COMPLETE 16:05

Digital Signature #
B9CE16CEB4F7977C
