
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000232
13:33 09/07/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:59
PRESSURIZING 3:03
STERILIZING 4:08
136.4 C 220kPa 4:08
136.4 C 225kPa 4:38
136.4 C 222kPa 5:08
136.4 C 223kPa 5:38
136.3 C 221kPa 6:08
136.3 C 221kPa 6:38
136.3 C 223kPa 7:08
136.4 C 223kPa 7:38
136.4 C 223kPa 7:38
VENTING 7:38
AIR DRYING 8:28
CYCLE COMPLETE 8:35

Digital Signature #
64326AB435D640F2
