

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3786

Name: RITTER2  
Sterilizer ID: V2313228  
07/16/2024 09:07 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.1 F 0.0  
2:00 72.7 F 0.0  
4:00 94.6 F 0.0  
6:00 91.6 F 0.1  
8:00 142.1 F 0.2  
10:00 162.0 F 2.6  
12:00 179.1 F 5.6  
14:00 201.0 F 8.9  
16:00 214.6 F 12.4  
18:00 220.9 F 6.7  
20:00 231.9 F 11.0  
22:00 230.4 F 7.5  
24:00 242.9 F 13.3  
26:00 242.3 F 12.0  
28:00 253.9 F 18.0  
30:00 264.3 F 24.2  
32:00 271.8 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.1  
1:00 272.7 F 29.8  
2:00 271.5 F 29.0  
3:00 272.6 F 29.9  
4:00 272.1 F 29.3  
  
Min 271.2 F 28.9  
Max 272.7 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.3 F 0.0  
5:00 207.3 F 0.0  
10:00 200.7 F 0.0  
15:00 185.1 F 0.0  
20:00 181.3 F 0.0  
25:00 179.8 F 0.0  
30:00 182.3 F 0.0  
  
Drying Complete

FILLING: 0:51  
HEATING: 32:09  
STERILIZING: 4:00  
VENTING: 3:34  
DRYING: 30:00  
TOTAL CYCLE: 01:10:34  
CYCLE COMPLETE

07/16/2024 10:17 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_