
Midmark M11 - v1.0.0
Total Cycles: 709

Name: MIDMARK
Sterilizer ID: V2178198
02/26/2021 02:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 117.7 F 0.0
2:00 117.9 F 0.1
4:00 142.8 F 1.5
6:00 190.7 F 4.9
8:00 210.3 F 10.0
10:00 223.8 F 5.3
12:00 228.1 F 9.1
14:00 236.7 F 9.6
16:00 240.8 F 11.6
18:00 238.4 F 9.4
20:00 252.1 F 16.2
22:00 263.8 F 23.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.1 F 28.5
2:00 272.5 F 29.0
3:00 271.3 F 28.1
4:00 272.4 F 28.7

Min 271.3 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.1 F 0.0
5:00 208.2 F 0.0
10:00 204.6 F 0.0
15:00 236.8 F 0.0
20:00 234.6 F 0.0
25:00 227.7 F 0.0
30:00 214.3 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 23:59
STERILIZING: 4:00
VENTING: 2:57
DRYING: 30:00
TOTAL CYCLE: 01:02:07
CYCLE COMPLETE

02/26/2021 03:05 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____