

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000436  
9:47 28/01/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:59  
PRESSURIZING 2:04  
STERILIZING 3:07  
136.0 C 229kPa 3:07  
136.0 C 230kPa 3:37  
136.4 C 230kPa 4:07  
136.3 C 233kPa 4:37  
136.4 C 233kPa 5:07  
136.3 C 232kPa 5:37  
136.0 C 230kPa 6:07  
136.4 C 233kPa 6:37  
136.4 C 233kPa 6:37  
VENTING 6:38  
AIR DRYING 7:29  
CYCLE COMPLETE 7:31  
-----

Digital Signature #  
811E9BA4E9BA77E6  
-----