
Midmark M11 - v1.0.5
Total Cycles: 698

Name: RITTER2
Sterilizer ID: V2313228
01/13/2022 03:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.6 F 0.0
2:00 177.8 F 3.7
4:00 201.7 F 8.9
6:00 224.2 F 8.1
8:00 233.8 F 10.9
10:00 235.4 F 9.1
12:00 238.2 F 9.7
14:00 249.4 F 15.3
16:00 261.6 F 22.4
18:00 272.6 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.5
1:00 272.0 F 29.4
2:00 272.6 F 29.6
3:00 272.6 F 29.8
4:00 272.9 F 30.1

Min 271.8 F 29.4
Max 273.0 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.9 F 0.0
5:00 225.0 F 0.0
10:00 220.1 F 0.0
15:00 200.8 F 0.0
20:00 192.3 F 0.0
25:00 200.9 F 0.0
30:00 192.9 F 0.0
Drying Complete

FILLING: 2:25
HEATING: 18:55
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:29
CYCLE COMPLETE

01/13/2022 04:37 PM
ACCEPTED: X
REJECTED:
Operator: U02