
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001717
12:42 21/03/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:41
PRESSURIZING 2:48
STERILIZING 3:52
136.4 C 223kPa 3:52
136.2 C 224kPa 4:22
136.4 C 222kPa 4:52
136.2 C 220kPa 5:22
136.4 C 224kPa 5:52
136.2 C 221kPa 6:22
136.4 C 224kPa 6:52
136.3 C 224kPa 7:22
136.3 C 224kPa 7:22
VENTING 7:22
AIR DRYING 8:13
CYCLE COMPLETE 8:20

Digital Signature #
EB81DAAA8885E87C
