
Midmark M11 - v1.0.0
Total Cycles: 857

Name: RITTER 1
Sterilizer ID: V2178198
06/01/2021 08:31 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 149.1 F 0.0
2:00 148.4 F 0.9
4:00 150.9 F 3.0
6:00 174.5 F 6.2
8:00 201.3 F 10.4
10:00 217.4 F 5.8
12:00 224.1 F 9.0
14:00 237.0 F 12.5
16:00 237.7 F 10.5
18:00 235.4 F 8.1
20:00 248.2 F 14.5
22:00 259.6 F 20.7
24:00 270.0 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.5 F 29.1
2:00 271.4 F 28.3
3:00 272.0 F 29.1
4:00 272.2 F 28.8

Min 271.1 F 28.3
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.6 F 0.0
5:00 217.0 F 0.0
10:00 217.7 F 0.0
15:00 218.9 F 0.0
20:00 238.1 F 0.0
25:00 228.9 F 0.0
30:00 230.1 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 24:58
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:03:18
CYCLE COMPLETE

06/01/2021 09:34 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____