
Midmark M11 - v1.0.0
Total Cycles: 4426

Name: RITTER 1
Sterilizer ID: V2178198
04/04/2024 08:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.3 F 0.0
2:00 81.4 F 0.3
4:00 82.6 F 1.2
6:00 118.6 F 3.4
8:00 148.0 F 6.4
10:00 169.2 F 10.0
12:00 187.9 F 9.0
14:00 211.8 F 4.0
16:00 224.9 F 7.0
18:00 234.0 F 10.5
20:00 233.9 F 8.2
22:00 238.7 F 10.9
24:00 236.4 F 8.1
26:00 248.5 F 14.6
28:00 259.7 F 20.8
30:00 270.0 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.4
1:00 272.6 F 28.8
2:00 272.3 F 29.0
3:00 272.0 F 28.4
4:00 272.8 F 29.2

Min 271.6 F 28.3
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.6 F 0.0
5:00 213.7 F 0.0
10:00 203.2 F 0.0
15:00 215.8 F 0.0
20:00 221.1 F 0.0
25:00 224.3 F 0.0
30:00 227.2 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 30:55
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 01:09:15
CYCLE COMPLETE

04/04/2024 09:52 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____