
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006766
13:34 19/12/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:22
PRESSURIZING 3:23
STERILIZING 4:25
136.5 C 225kPa 4:25
136.4 C 225kPa 4:55
136.4 C 220kPa 5:25
136.4 C 224kPa 5:55
136.4 C 224kPa 6:25
136.2 C 224kPa 6:55
136.3 C 224kPa 7:25
136.4 C 225kPa 7:55
136.4 C 225kPa 7:55
VENTING 7:55
AIR DRYING 8:45
CYCLE COMPLETE 8:50

Digital Signature #
67D970CD875C9882
