
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
0.9uS / 0.5ppm
CYCLE NUMBER 001453
13:41 09/27/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:56
PRESSURIZING 6:46
STERILIZING 8:14
136.0 C 220kPa 8:14
136.3 C 219kPa 8:44
136.5 C 223kPa 9:14
136.4 C 223kPa 9:44
136.1 C 221kPa 10:14
136.4 C 220kPa 10:44
136.4 C 222kPa 11:15
136.3 C 222kPa 11:44
VENTING 11:44
AIR DRYING 12:46
CYCLE COMPLETE 32:06

Digital Signature #
F7C0F6AE54FC3557
