
Midmark M11 - v1.0.5
Total Cycles: 3206

Name: RITTER2
Sterilizer ID: V2313228
12/11/2023 04:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.2 F 0.0
2:00 178.5 F 3.5
4:00 201.8 F 9.0
6:00 227.9 F 8.7
8:00 233.2 F 10.6
10:00 232.7 F 8.4
12:00 239.9 F 10.4
14:00 248.6 F 14.8
16:00 261.6 F 22.2
18:00 272.7 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.3
1:00 272.6 F 29.4
2:00 272.7 F 29.8
3:00 271.9 F 29.7
4:00 272.2 F 29.4

Min 271.5 F 29.2
Max 273.1 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.4 F 0.0
5:00 226.0 F 0.0
10:00 214.4 F 0.0
15:00 202.2 F 0.0
20:00 185.6 F 0.0
25:00 189.4 F 0.0
30:00 201.3 F 0.0
Drying Complete

FILLING: 1:56
HEATING: 18:54
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 00:58:15
CYCLE COMPLETE

12/11/2023 05:34 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____