
Midmark M11 - v1.0.0
Total Cycles: 5329

Name: RITTER 1
Sterilizer ID: V2178198
02/10/2025 02:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.5 F 0.0
2:00 155.6 F 3.3
4:00 178.2 F 7.1
6:00 201.8 F 11.7
8:00 217.0 F 5.8
10:00 228.1 F 11.1
12:00 230.7 F 7.5
14:00 244.3 F 13.5
16:00 242.5 F 11.7
18:00 254.6 F 18.0
20:00 265.9 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.3
1:00 273.1 F 29.3
2:00 271.8 F 29.0
3:00 271.9 F 28.8
4:00 272.4 F 29.3

Min 270.9 F 28.2
Max 273.2 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.3 F 0.0
5:00 207.1 F 0.0
10:00 235.3 F 0.0
15:00 234.8 F 0.0
20:00 228.5 F 0.0
25:00 228.5 F 0.0
30:00 226.7 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 21:39
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:00:24
CYCLE COMPLETE

02/10/2025 03:13 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____