
Midmark M11 - v1.0.5
Total Cycles: 123

Name: RITTER2
Sterilizer ID: V2313228
06/21/2021 12:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 102.2 F 0.0
2:00 106.9 F 0.9
4:00 124.7 F 3.4
6:00 154.3 F 6.8
8:00 181.2 F 10.9
10:00 200.4 F 4.9
12:00 217.1 F 7.3
14:00 227.2 F 11.9
16:00 231.2 F 8.8
18:00 240.1 F 11.5
20:00 243.7 F 13.0
22:00 255.5 F 19.2
24:00 266.1 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.5
1:00 272.9 F 30.1
2:00 271.6 F 29.4
3:00 272.8 F 30.3
4:00 272.2 F 29.7

Min 271.5 F 29.4
Max 273.0 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.8 F 0.0
5:00 200.6 F 0.0
10:00 209.8 F 0.0
15:00 190.9 F 0.0
20:00 176.8 F 0.0
25:00 176.5 F 0.0
30:00 163.4 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 25:44
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:04:02
CYCLE COMPLETE

06/21/2021 01:24 PM
ACCEPTED: X
REJECTED:
Operator: U01