
Midmark M11 - v1.0.0
Total Cycles: 3266

Name: RITTER 1
Sterilizer ID: V2178198
05/23/2023 09:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 205.7 F 0.0
2:00 198.7 F 0.4
4:00 212.2 F 5.8
6:00 221.9 F 11.4
8:00 228.5 F 7.3
10:00 242.1 F 13.2
12:00 241.8 F 11.3
14:00 240.1 F 10.2
16:00 254.3 F 17.3
18:00 266.2 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.0
1:00 272.6 F 28.8
2:00 272.0 F 28.2
3:00 271.4 F 28.1
4:00 272.6 F 28.8

Min 271.3 F 27.8
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.8 F 0.0
5:00 230.7 F 0.0
10:00 240.1 F 0.0
15:00 234.7 F 0.0
20:00 227.0 F 0.0
25:00 226.8 F 0.0
30:00 226.9 F 0.0
Drying Complete

FILLING: 1:49
HEATING: 19:37
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:38
CYCLE COMPLETE

05/23/2023 10:49 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____