

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2527

Name: RITTER 1  
Sterilizer ID: V2178198  
11/08/2022 10:02 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 143.1 F 0.0  
2:00 153.8 F 0.1  
4:00 173.4 F 3.0  
6:00 185.5 F 6.9  
8:00 204.3 F 11.5  
10:00 219.6 F 5.2  
12:00 228.8 F 10.2  
14:00 236.4 F 8.0  
16:00 242.0 F 12.3  
18:00 240.5 F 10.3  
20:00 253.2 F 16.7  
22:00 264.2 F 23.3  
24:00 271.3 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.2  
1:00 272.1 F 28.7  
2:00 271.3 F 28.3  
3:00 272.1 F 28.7  
4:00 271.2 F 28.1

Min 271.2 F 28.1  
Max 272.2 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.0 F 0.0  
5:00 209.2 F 0.0  
10:00 220.4 F 0.0  
15:00 223.2 F 0.0  
20:00 229.7 F 0.0  
25:00 233.2 F 0.0  
30:00 231.6 F 0.0  
Drying Complete

FILLING: 0:47  
HEATING: 24:00  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:01:57  
CYCLE COMPLETE

11/08/2022 11:03 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_