

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1242

Name: RITTER2  
Sterilizer ID: V2313228  
06/29/2022 01:17 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 162.6 F 0.0  
2:00 163.6 F 3.5  
4:00 177.1 F 7.4  
6:00 205.6 F 12.2  
8:00 216.6 F 6.7  
10:00 230.8 F 11.9  
12:00 232.3 F 8.9  
14:00 238.6 F 10.2  
16:00 245.7 F 13.8  
18:00 258.0 F 20.3  
20:00 269.1 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.2  
1:00 272.5 F 29.8  
2:00 271.5 F 29.1  
3:00 272.1 F 29.8  
4:00 272.5 F 29.7  
  
Min 271.3 F 29.0  
Max 272.6 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.5 F 0.0  
5:00 221.4 F 0.0  
10:00 223.2 F 0.0  
15:00 225.0 F 0.0  
20:00 225.3 F 0.0  
25:00 225.8 F 0.0  
30:00 225.6 F 0.0  
Drying Complete

FILLING: 1:11  
HEATING: 21:07  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:59:26  
CYCLE COMPLETE

06/29/2022 02:16 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_