
Midmark M11 - v1.0.0
Total Cycles: 2873

Name: RITTER 1
Sterilizer ID: V2178198
02/08/2023 09:57 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.7 F 0.0
2:00 164.8 F 2.4
4:00 169.7 F 6.1
6:00 191.5 F 10.6
8:00 212.6 F 4.5
10:00 225.3 F 9.3
12:00 235.1 F 9.3
14:00 239.7 F 11.2
16:00 236.8 F 9.0
18:00 250.5 F 15.7
20:00 262.1 F 22.3
22:00 272.7 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.7 F 29.3
2:00 271.9 F 28.7
3:00 271.5 F 28.7
4:00 272.6 F 29.3

Min 271.3 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.6 F 0.0
5:00 224.7 F 0.0
10:00 230.7 F 0.0
15:00 231.0 F 0.0
20:00 220.4 F 0.0
25:00 214.9 F 0.0
30:00 213.7 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 22:26
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:00:39
CYCLE COMPLETE

02/08/2023 10:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____