

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000700  
7:22 28/10/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:01  
PRESSURIZING 3:02  
STERILIZING 4:07  
136.0 C 220kPa 4:07  
135.9 C 221kPa 4:37  
135.8 C 220kPa 5:07  
136.2 C 222kPa 5:37  
136.1 C 221kPa 6:07  
136.1 C 222kPa 6:37  
136.0 C 220kPa 7:07  
136.1 C 221kPa 7:37  
136.1 C 221kPa 7:37  
VENTING 7:37  
AIR DRYING 8:28  
CYCLE COMPLETE 8:38

-----  
Digital Signature #  
C370AF8D2FB35F82  
-----