
Midmark M11 - v1.0.0
Total Cycles: 3817

Name: RITTER 1
Sterilizer ID: V2178198
09/26/2023 09:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.3 F 0.0
2:00 172.9 F 0.1
4:00 181.9 F 1.7
6:00 209.6 F 5.4
8:00 219.8 F 9.7
10:00 231.3 F 9.1
12:00 235.4 F 9.7
14:00 237.2 F 9.3
16:00 245.0 F 12.7
18:00 242.7 F 11.3
20:00 254.9 F 17.5
22:00 265.5 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 27.9
1:00 272.5 F 28.6
2:00 271.5 F 27.9
3:00 271.8 F 28.5
4:00 272.5 F 28.4

Min 271.0 F 27.7
Max 272.5 F 28.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.4 F 0.0
5:00 206.7 F 0.0
10:00 190.5 F 0.0
15:00 187.1 F 0.0
20:00 183.1 F 0.0
25:00 181.1 F 0.0
30:00 180.1 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 23:52
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:02:30
CYCLE COMPLETE

09/26/2023 10:10 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____