

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002375  
11:31 13/12/2024  
AIR DRYING ONLY  
60 min. MAXIMUM  
CYCLE START 0:00  
AIR DRYING 0:12  
CYCLE COMPLETE 0:42  
-----

Digital Signature #  
546B47914D24E0BE  
-----