
Midmark M11 - v1.0.5
Total Cycles: 2726

Name: RITTER2
Sterilizer ID: V2313228
08/02/2023 02:47 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 97.9 F 0.0
2:00 99.4 F 0.0
4:00 114.2 F 0.0
6:00 153.5 F 0.0
8:00 176.8 F 0.0
10:00 200.0 F 1.7
12:00 210.8 F 5.0
14:00 221.2 F 8.2
16:00 230.9 F 11.9
18:00 231.0 F 7.8
20:00 241.6 F 12.6
22:00 239.2 F 10.4
24:00 237.1 F 8.7
26:00 249.2 F 15.0
28:00 260.0 F 21.0
30:00 269.7 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.7 F 29.2
2:00 271.7 F 28.4
3:00 272.7 F 29.4
4:00 272.1 F 28.7

Min 271.6 F 28.4
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.4 F 0.0
5:00 196.9 F 0.0
10:00 199.0 F 0.0
15:00 197.1 F 0.0
20:00 195.2 F 0.0
25:00 195.5 F 0.0
30:00 200.9 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 31:02
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:09:09
CYCLE COMPLETE

08/02/2023 03:56 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____