

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3819

Name: RITTER 1  
Sterilizer ID: V2178198  
09/26/2023 11:14 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 180.6 F 0.0  
2:00 180.5 F 2.8  
4:00 208.4 F 7.9  
6:00 226.1 F 12.1  
8:00 229.9 F 9.1  
10:00 237.2 F 9.8  
12:00 242.5 F 12.1  
14:00 241.3 F 10.7  
16:00 254.2 F 17.2  
18:00 265.3 F 23.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.0  
1:00 272.2 F 28.4  
2:00 272.0 F 28.6  
3:00 271.8 F 28.1  
4:00 272.5 F 28.8  
  
Min 271.2 F 27.9  
Max 272.6 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.1 F 0.0  
5:00 220.3 F 0.0  
10:00 220.2 F 0.0  
15:00 218.8 F 0.0  
20:00 216.4 F 0.0  
25:00 212.9 F 0.0  
30:00 206.6 F 0.0  
Drying Complete

FILLING: 1:41  
HEATING: 19:46  
STERILIZING: 4:00  
VENTING: 3:27  
DRYING: 30:00  
TOTAL CYCLE: 00:58:53  
CYCLE COMPLETE

09/26/2023 12:12 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_