
Midmark M11 - v1.0.5
Total Cycles: 2117

Name: RITTER2
Sterilizer ID: V2313228
03/03/2023 11:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	101.1 F	0.0
2:00	112.0 F	0.0
4:00	143.9 F	0.1
6:00	148.8 F	3.7
8:00	183.0 F	8.6
10:00	201.1 F	8.0
12:00	223.1 F	7.7
14:00	237.2 F	13.2
16:00	239.6 F	10.9
18:00	238.2 F	9.7
20:00	253.3 F	17.1
22:00	265.9 F	24.7

C103: HEATUP MODE
STOP PRESSED
CHAMBER TEMP 253F
CHAMBER PRESS 17.0PSI
** ITEMS NOT STERILE **

03/03/2023 12:21 PM