
Midmark M11 - v1.0.0
Total Cycles: 2536

Name: RITTER 1
Sterilizer ID: V2178198
11/09/2022 10:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.1 F 0.0
2:00 187.0 F 3.6
4:00 194.9 F 8.0
6:00 216.8 F 11.4
8:00 226.0 F 7.8
10:00 238.0 F 13.2
12:00 238.6 F 10.4
14:00 238.0 F 8.3
16:00 250.6 F 15.4
18:00 262.5 F 22.3
20:00 272.6 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.5 F 29.0
2:00 272.1 F 28.4
3:00 271.4 F 28.2
4:00 272.5 F 29.0

Min 271.4 F 28.0
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.7 F 0.0
5:00 224.4 F 0.0
10:00 230.9 F 0.0
15:00 228.9 F 0.0
20:00 227.6 F 0.0
25:00 224.8 F 0.0
30:00 222.1 F 0.0
Drying Complete

FILLING: 1:28
HEATING: 20:18
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:56
CYCLE COMPLETE

11/09/2022 11:56 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____