
Midmark M11 - v1.0.5
Total Cycles: 1127

Name: RITTER2
Sterilizer ID: V2313228
05/27/2022 08:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 146.8 F 0.0
2:00 144.7 F 1.5
4:00 149.4 F 4.0
6:00 177.6 F 8.1
8:00 199.4 F 13.0
10:00 218.0 F 7.0
12:00 231.0 F 12.0
14:00 232.2 F 8.9
16:00 239.0 F 10.6
18:00 244.8 F 13.4
20:00 256.8 F 19.7
22:00 267.6 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.3
1:00 272.5 F 30.0
2:00 271.7 F 29.2
3:00 272.0 F 29.9
4:00 272.5 F 29.7

Min 271.3 F 29.1
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.7 F 0.0
5:00 207.7 F 0.0
10:00 223.2 F 0.0
15:00 218.2 F 0.0
20:00 215.8 F 0.0
25:00 216.2 F 0.0
30:00 215.5 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 23:25
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:01:28
CYCLE COMPLETE

05/27/2022 09:43 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____