
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007800
8:41 27/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 0:52
PRESSURIZING 0:56
STERILIZING 1:58
136.4 C 220kPa 1:58
135.9 C 219kPa 2:28
136.2 C 220kPa 2:58
136.1 C 217kPa 3:28
136.4 C 221kPa 3:58
136.4 C 222kPa 4:28
136.2 C 220kPa 4:58
136.4 C 223kPa 5:28
136.4 C 223kPa 5:28
VENTING 5:28
AIR DRYING 6:16
CYCLE COMPLETE 7:26

Digital Signature #
AB1EA8ABB55D7A3E
