
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000170
9:58 16/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:28
PRESSURIZING 4:28
STERILIZING 5:42
136.0 C 222kPa 5:42
136.4 C 225kPa 6:12
136.4 C 225kPa 6:42
136.3 C 223kPa 7:12
136.3 C 224kPa 7:42
136.3 C 225kPa 8:12
136.3 C 225kPa 8:42
136.3 C 225kPa 9:12
136.3 C 225kPa 9:12
VENTING 9:12
AIR DRYING 10:12
CYCLE COMPLETE 25:13

Digital Signature #
32D4C5ABC49AB24E
